

Description

The TAR DNA binding protein of 43 kDa (TDP-43 or TARDBP) is a highly conserved and ubiquitously expressed nuclear protein with roles in transcription and splicing regulation. It is also the major component of ubiquitinpositive cytoplasmic inclusions found in the brains of patients with frontotemporal lobar degeneration (FTLD) and amyotrophic lateral sclerosis (ALS). In addition, TDP-43-containing aggregates are found in a significant number of patients with Alzheimer's Disease (AD) and other neuromuscular disorders. The majority of TDP-43 protein found in cytoplasmic inclusions is truncated, and it has been shown that the C-terminal domain is intrinsically prone to aggregation. Mutations in the Cterminal region of the TDP-43 gene have been associated with both ALS and FTLD, and are thought to facilitate ubiquitination and phosphorylation of the TDP-43 protein, leading to the formation of pathological inclusions and eventual neurodegeneration. Analysis of TDP-43 levels in plasma may allow the indexing of TDP-43 pathology within the brain to aid in the diagnosis of different forms of dementia and distinguish between TDP-43 proteinopathy and tauopathy. The Simoa TDP-43 assay has been developed with a full-length protein calibrator and antibodies against AA 203 - 209 and the C-terminal region; it is expected to detect both full-length and pathological, truncated forms of the protein.

Calibration Curve: Calibrator concentrations and Lower Limit of Quantification depicted.



Lower Limit of Quantification (LLOQ): Triplicate measurements of serially diluted calibrator were read back on the calibration curve over 12 runs each for 2 reagent lot across 2 instruments (12 runs total).

Limit of Detection (LOD): Calculated as 2.5 standard deviations from the mean of background signal read back on each calibration curve over 12 runs each for 2 reagent lot across 2 instruments (12 runs total).

| Analytical LLOQ | 8.23 pg/mL pooled CV 13% mean recovery 113% | |
|-------------------------------------|----------------------------------------------------------|--|
| LOD | 2.48 pg/mL range 0.234-5.08 pg/mL | |
| Dynamic range (serum and plasma) | 0-8000 pg/mL | |
| Diluted Sample volume* | 100 μL per measurement | |
| Tests per kit | 96 | |
| *See Kit Instruction for details | | |

Endogenous Sample Reading: Healthy donor matched EDTA plasma (n=20), and serum (n=20) were measured. Nine CSF samples were measured. Bars depict median with interquartile range. Orange line represents functional LLOQ.



Quanterix Corporation 900 Middlesex Turnpike, Billerica, MA 01821 techsupport@quanterix.com ©2018 Quanterix, Inc. Simoa[®] and Quanterix[®] are registered trademarks of Quanterix, Inc. For research use only. Not for use in diagnostic procedures. DS-0134 05 Page 1 of 2

| Simoa® | TDP-43 Adva | intage Kit |
|----------|---------------|-------------|
| HD-1/HD- | -X Data Sheet | Item 103293 |

| Sample Type | Mean TDP-43 pg/mL | Median TDP-43 pg/mL | % Above LOD |
|----------------|----------------------|------------------------|----------------|
| Serum | 159 | 130 | 100% |
| Plasma | 209 | 214 | 100% |
| CSF | 127 | 25.3 | 69% |

Precision: Measurements of 1 plasma-based panel, 3 serum-based panels and 2 calibrator-based controls. Triplicate measurements were made for 6 runs each for 2 reagent lots across 2 instruments (12 runs total, 36 measurements).

| Sample | Mean (pg/mL) | Within run CV | Btwn run CV | Btwn inst CV | Btwn lot CV |
|-----------|-----------------|------------------|----------------|-----------------|----------------|
| Control 1 | 195 | 4.3% | 4.0% | 3.4% | 1.8% |
| Control 2 | 1953 | 3.1% | 4.5% | 2.5% | 2.5% |
| Panel 1 | 59.0 | 11.8% | 11.3% | 0.1% | 1.1% |
| Panel 2 | 73.5 | 6.2% | 14.1% | 3.3% | 1.3% |
| Panel 3 | 316 | 3.2% | 11.1% | 7.3% | 3.4% |
| Panel 4 | 1771 | 4.0% | 8.2% | 5.7% | 6.0% |

Spike and Recovery: 2 serum, 2 EDTA plasma and 2 CSF samples were spiked at high and low concentrations within the range of the assay and analyzed on HD-1.

Observed recovery was consistently low in serum and plasma, but results from dilutional linearity and immunodepletion experiments support specificity of the assay signal.

Dilution Linearity: 2 spiked EDTA plasma and 2 spiked serum samples were diluted 2x serially with Sample Diluent from MRD (4x) to 32x. 2 spiked CSF samples were diluted 2x serially with Sample Diluent from MRD (4x) to 64x. All dilutions were performed offline with Sample Diluent and run neat on the HD-1 analyzer. Offline dilution is recommended when assessing dilution linearity for this assay on HD-1/HD-X.

| Spike and Recovery | Mean = 32% |
|---------------------------|-----------------|
| (Serum/Plasma) | Range: 26-42% |
| Dilution Linearity | Mean = 116% |
| (Serum/Plasma) (32x) | Range: 101–135% |
| Spike and Recovery | Mean = 92% |
| (CSF) | Range: 87-100% |
| Dilution Linearity | Mean = 103% |
| (CSF) (64x) | Range: 98–109% |

Immuno-depletion: 1 Serum and 2 plasma samples were separately incubated with TDP-43 beads and Antibody isotype control beads prior to analysis on HD-1. Mean depletion was 98%.

The Simoa TDP-43 Advantage assay kit is formulated for use on the SR-X[®], HD-1, or HD-X[®] platform. Data in this document was obtained from runs on the HD-1 platform unless otherwise noted. Some differences in performance claims between SR-X and HD-1/HD-X may be observed when comparing datasheets for these platforms. This may be due to experiments run at different time-points with different reagent lots and different samples, or it may be due to minor differences in antibody and analyte behavior in the different assay formats.

Quanterix Corporation 900 Middlesex Turnpike, Billerica, MA 01821 techsupport@quanterix.com ©2018 Quanterix, Inc. Simoa[®] and Quanterix[®] are registered trademarks of Quanterix, Inc. For research use only. Not for use in diagnostic procedures. DS-0134 05 Page 2 of 2