Quanterix
The Science of Precision Health

# Simoa® Qualification Test (SQT) V2 Kit
## Certificate of Analysis

| | |
|---|---|
| Product Number: | 103591 |
| Lot Number: | 503358 |
| Expiration: | 25-Sep-2022 |

| Component | Part Number | Lot Number |
|---|---|---|
| Bead Reagent | 103581 | 126804 |
| Detector Reagent | 103583 | 126805 |
| SBG Reagent | 103585 | 126806 |
| Control 1 | 103589 | 127002 |
| Control 2 | 103590 | 127002 |
| RGP Reagent | 103159 | N/A[1] |

[1] RGP is not Kit Lot Specific

| | |
|---|---|
| Bead Count (beads/mL) | $2.00 \times 10^7$ |
| % Bead Monomeric | 94% |
| Isingle | 167 |

| | | | |
|---|---|---|---|
| **Review/Approval** | Shivani Goel | QC, Manager | Shivani Goel 17Feb2022 |
| | **Name** | **Title** | **Signature/Date** |

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, MA 01821
techsupport@quanterix.com

©2019 Quanterix, Inc. Simoa™ is a registered trademark of Quanterix, Inc
For research use only. Not for use in diagnostic procedures.