Quanterix
The Science of Precision Health

Simoa® Discovery SBG Diluent
Certificate of Test

Product Number:	101956
Lot Number:	306517
Expiration:	6-Mar-2024
Storage:	2-8°C

Laboratory Analysis

Characteristic	Result	pH Target
pH Measurement	PASS	7.35 - 7.45

Review/Approval Sonia Khillan (Name) | QC Manager (Title) | [signature] 20MAR2023 (Signature/Date)

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, MA 01821
techsupport@quanterix.com

©Quanterix Corp. Simoa® is a registered trademark of Quanterix Corp.
For research use only. Not for use in diagnositic procedures