Quanterix®
Discovery Fueled by Ultra-Sensitivity

Simoa® Homebrew Bead Diluent (500 mL)

Certificate of Test

Product Number:	101362
Lot Number:	403611
Expiration:	5-Aug-2025
Storage:	2-8°C

Laboratory Analysis

Characteristic	pH Target	PASS / FAIL
pH Measurement	7.75 - 7.85	PASS

RESULT

pH Reading	7.75

	Name	Title	Signature/Date
Review/Approval	Heidi Caswell	Director, QC	Heidi Caswell 14 Mar 2024

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, 01821
techsupport@quanterix.com

©2023 Quanterix Corp. Simoa® is a registered trademark of Quanterix Corp.

For research use only. Not for use in diagnostic procedures.